## EUROMET PROJECT FINAL REPORT



7.	Progress cont: All results agree well within the stated uncertainties. CEM's result for the 1000 mm gauge block is considerably larger than the results from SP and NMi, but since their result for the 500 mm block agree very well, there does not seem to be any significant systematic error present in their equipment. One of the faces of the 250 mm gauge block was concave, which made the results differ somewhat depending on which face of the block that was wrung. Nevertheless the mean values from SP and NRC show good agreement.				
8.	Coordinator's name: SP-Swedish National Testing and Research Institute, Optics and Geometry Stefan Källberg Address: Brinellgatan 4, S-501 15 Borås SWEDEN				
	Telephone: +46 33 16 56 26 Telefax: +46 33 16 56 20 E-mail: stefan.kallberg@sp.se				
9. Completion date: January 2001		10. Coordinator's signature:		11. Date:	