

EUROMET PROJECT FINAL REPORT

1. Ref. No.: 653	2. Subject Field: THERMOMETRY	
3. Type of collaboration: Cooperation		
4A. Partners: BEV, PTB (institutions)	4B. CEC funded? NO	
5. Participating countries: AT, DE		
6. Title: Traceability of temperature		
<p>7. Progress:</p> <p>In January 2002 a SPRT from the BEV was calibrated at the PTB in the temperature range:</p> <p style="text-align: center;">$83,8054\text{ K} \leq T \leq 234,3156\text{ K}$</p> <p>In this range the BEV is traceable to the PTB therefore. All measurements are finished.</p>		
<p>8. Coordinator's name: Prof. Dr. Franz Adunka</p> <p>Address: Bundesamt für Eich- und Vermessungswesen (BEV), Arltgasse 35, A-1163 Vienna, Austria</p> <p>Telephone: ++43-1-49110-537 Telefax: ++43-1-49 20 875 E-mail: f.adunka@metrologie.at</p>		
9. Completion date: January 2002	8. Coordinator's signature:	11. Date: April 3rd, 2002