

## Publishable Summary for 22IEM06 S-CALe Up

### Self-calibrating photodiodes for UV and exploitation of induced junction technology

#### Overview

Photonics and optics are exciting, rapidly evolving technologies of great value (over 100 billion € in EU) and of importance across industry, environment, health, medicine, energy, lighting and science. The photonic industry and standardisation organisations request miniaturised, cost-effective, integrated and self-calibrating measurement systems that cannot be provided by traditional methods. Recent developments exploiting predictable photodiodes in various ways have demonstrated improved uncertainty of responsivity to 10 ppm and the proof-of-concept of an NMI-on-a-chip suitable for miniaturisation. This project will demonstrate exploitation of the new photodiodes as built-in standards in various applications and develop improved standard detectors for the UV range by exploiting new methods.

#### Need

There is rapid and exciting technology development within photonics, which is one of six EU defined Key Enabling Technologies (KET) used in climate monitoring, medical treatment, health and photonic industries, energy saving illumination by light-emitting diodes (LEDs), electricity production, science, and many more applications. The technological development trend is towards miniaturisation, more integrated measurement systems and distribution of standalone sensor systems in possibly remote locations. Current metrological systems are not capable of calibrating photodiodes in integrated systems nor in remote locations. Therefore, the European technology platforms Photonics21 and Quantum Flagship highlight the integration of self-calibrating systems and products as one of the technology, research, and innovation challenges ahead.

Previous iMERA+, EMRP and EMPIR projects have developed the Predictable Quantum Efficient Detector (PQED), which has proven to have an extremely low external quantum deficiency (EQD) of around 10 ppm and an undetectable drift over 10 years. These properties make the PQED a very attractive calibration standard detector that complies well with the low-cost and high accuracy transfer standard requested by CIPM's Consultative Committee for Photometry and Radiometry (CCPR).

Demonstration of the utility and exploitability of the PQED's properties in various applications is necessary to increase their technological readiness level, bringing them closer to uptake by relevant industry and end-users.

The ultraviolet (UV) range is of great importance to health and earth observations but suffers from detectors with poor stability and lack of predictability. The International Commission on Illumination (CIE), as the world standardisation organisation for light and lighting, has requested research on new low-cost, high accuracy, primary standard detectors of optical radiation allowing better traceability from 200 nm – 2000 nm, which is of importance to industry and will also cover the UV range.

Improved understanding of the measurement system provided by advanced modelling of photodiodes supported by characterisation measurements will be key in enabling better traceability and will trigger new measurement techniques in remote and integrated measurement systems. In a similar way, dual-mode detectors are needed to enable a pure experimental self-calibrating measurement system in a wider range of applications for future exploitation also outside the spectral range covered by silicon technology where photodiodes cannot be accurately modelled.

**Report Status:**  
**PU – Public, fully open**

**Publishable Summary**

Issued: June 2023

*Funded by the European Union. Views and opinions expressed are however those of the author(s) only and do not necessarily reflect those of the European Union or EURAMET. Neither the European Union nor the granting authority can be held responsible for them.*

*The project has received funding from the European Partnership on Metrology, co-financed from the European Union's Horizon Europe Research and Innovation Programme and by the Participating States.*

European Partnership

Co-funded by the European Union

**METROLOGY PARTNERSHIP**



## Objectives

The overall objective is to demonstrate the use of improved PQED detectors and associated technology as an NMI-on-a-chip in various photonics applications and to develop models and improved detector manufacturing techniques in order to provide improved traceability into the more challenging UV and NIR spectral range.

The specific objectives of the project are:

1. To develop **3D simulation models** of photodiode charge carrier transportation in Predictable Quantum Efficient Detectors (PQED) for better physical representation, higher calculation speed, wider availability and improved uncertainty approaching 1 ppm between 500 nm and 900 nm, and to extend the quantum yield prediction from 400 nm down to 250 nm with an uncertainty better than 0.1 % and better than 0.2 % from 250 nm to 200 nm.
2. To **use PQEDs** with very low spectral responsivity uncertainty in the 400 nm to 850 nm range as built-in references **in different applications** (e.g. optical power measurement, fibre optics, pulsed laser radiation, photometry without  $V(\lambda)$  filters) taking into account practical aspects, such as current measurements, stray light, geometry, heat and dark current variations.
3. To develop and **fabricate improved photodiodes for the UV range** and validate their stability and suitability as a spectral responsivity standard from 400 nm to 200 nm. Additionally, to develop thermal simulations and packaging technology of dual-mode detectors with heat equivalence better than 0.03 % suitable for implementation at any wavelength over the spectral range from 200 nm to 1000 nm.
4. To **extend the spectral response range** of photodiodes between i) 200 nm and 400 nm, and ii) 850 nm and 1050 nm, with a target uncertainty better than 0.2 %. For this, the improved detectors and packaging developed in objective 3, and improved charge carrier simulation and quantum yield modelling developed in objective 1, will be used.
5. To demonstrate the establishment of **an integrated European metrology infrastructure and to facilitate the take up of the technology** and measurement infrastructure developed in the project by the measurement supply chain (accredited laboratories, instrument manufacturers), standards developing organisations (CIE), technical committees (e.g. EURAMET TC-PR, CCPR) and end users (photonics industry).

## Progress beyond the state of the art and results

### *Improved simulation models*

The project builds on the EMPIR 18SIB10 chipS-CALe project, where improved self-induced PQED trap detectors with a record low external quantum deficiency below 10 ppm were developed for the 400 nm to 850 nm spectral range. Simplified 3D simulation models were developed and new experimental techniques where simulation fitting to experimental photocurrent measurement at only one wavelength was proven to be sufficient to predict the responsivity from 400 nm to 850 nm. In previous projects, quantum yield (QY) was found significant at wavelengths as high as 470 nm and to increase with decreasing wavelength. However, the quantum yield was also found calculable to an uncertainty around 200 ppm down to 360 nm. Current 3D simulation capabilities are limited to one NMI, on one type of software, with known limiting simplifications due to calculation speed and the current TRL-level in the 400 nm – 850 nm spectral range is around 3 to 4.

In S-CALe Up simulation models will be used to

- I. improve predictability of the PQED to achieve an uncertainty that goes beyond the IQD losses,
- II. independently predict the spectral response of PQEDs from 300 nm to 1000 nm, the spectral range covered by the key comparison CCPR-K2.b.
- III. develop improved standard detectors and prediction methods designed for the UV spectral range from 200 nm to 400 nm, the spectral range covered by the key comparison CCPR-K2.c.

Furthermore, reflectance losses and quantum yield will be studied and modelled to enable the development of independently predictable standard detectors for the full spectral ranges covered by CCPR-K2.b and CCPR-K2.c.

### *Applications for photometry, fibre optics and PICs exploiting PQEDs*

The PQEDs have proven to be extremely stable with an undetectable drift, independent of wavelength, over ten years with a responsivity directly linked to fundamental constants. This makes the PQED an ideal transfer standard ready to be exploited in applications.

In S-CALe Up, three different applications for providing services to industry with the PQED as a built-in reference will be demonstrated, which will bring the technology closer to uptake by end-users. These demonstrators will simplify realisations, improve traceability and are expected to produce improved calibration and measurement capabilities (CMCs) with reduced uncertainty for the laboratories developing the demonstrators.

### *Develop and fabricate improved photodiodes for the UV range and packaging technology*

Silicon transfer standard pn-photodiodes for the UV spectral range from 200 nm to 400 nm are known to deteriorate under UV exposure, but new process advancements have proven to produce more stable photodiodes. Alternative materials and diode types such as PtSi and GaAsP are also used, but they suffer from a large non-uniformity. The self-induced photodiodes in the PQED may be less susceptible to radiation damage, but this has not been tested. Alternative calibration methods based on purely experimental self-calibrating techniques referred to as dual-mode detectors (DMD) are promising and generic methods to provide traceability to a wide spectral range. The techniques have been developed and demonstrated at room temperature with an uncertainty below 0.05 %. Current packaging technologies for dual-mode detectors are limited to exploiting silicon photodiodes, as they are heated through forward bias, and have a beam position dependent thermal non-equivalence around 280 ppm/mm.

In this project, improved photodiodes for the UV range with improved and validated stability will be developed taking advantage of the predictability of the photodiodes. Combining simulation models and fast material characterisations will enable many different processes to be tested to produce optimum and robust passivation. Furthermore, the influence and optimising of the current read-out contacts will be studied. Improved packaging suitable for future industrialisation and implementation with other types of photodiodes than silicon will be developed.

### *Extended and validated spectral response scales*

In this project, improved photodiodes and modelling will be exploited to establish and validate the fully predictable spectral response scales covering the CCPR-K2.c range from 200 nm to 400 nm and the CCPR-K2.b range from 300 nm to 1000 nm with a target uncertainty below 0.2 %. This will simplify the realisations of spectral response scale as fewer calibration points are needed, and will enable implementation of independent traceability techniques exploiting the state of the art PQEDs.

## **Outcomes and Impact**

### *Outcomes for industrial and other user communities*

Industry in general require improved and simplified measurement standards suitable for miniaturisation. The development trend is moving in the direction of more compact and integrated systems, and this project aims to enable future implementation of traceability to such systems where existing measurement techniques are not capable of providing traceability. By exploiting PQEDs in specific metrological applications, cost efficient realisations of primary photometers and power meters of importance to industry will be demonstrated. The outstanding self-calibrating technology developed in the previous project 18SIB10 chipS-CALe is expected to continue to develop towards more user-friendly solutions so that more applications can profit from it. Implementation of PQEDs in applications will bring the technology closer to the market.

### *Outcomes for the metrology and scientific communities*

Realising primary standard techniques based on silicon photodiodes outline new ways to provide traceability based on the built-in responsivity linked to fundamental constants. With the realisation of the units and their dissemination occurring through the same artefact, the PQED provides the means to shorten the traceability chain and improve the measurement uncertainty. The microelectronic processing makes the primary standard suitable for miniaturisation and integration into measurement systems where traceability was previously not achievable. Furthermore, the new techniques with predictable spectral response throughout the whole spectral

range of interest will ensure better global harmonisation of measurement results and improved access throughout the metrological community to the highest metrological standards.

The structure in the project will enable NMIs to develop and exploit their specific niche of expertise which will be used to meet the project's overall goals, and to develop European robustness in services where more than one NMI can provide the needed services to stakeholders. This collaboration will strengthen the metrological expertise of all contributing NMIs, including small and emerging institutes, support EURAMET's goal for smart specialisation, and assure an internationally leading position.

#### *Outcomes for relevant standards*

The International Commission of Illumination (CIE) is the international standardisation organisation on all aspects of light and lighting. Today's conventional techniques for realising absolute standard detectors require expensive equipment, high skill levels to operate, time-consuming methods and interpolation functions to establish a continuous spectral response scale.

The early outcomes of this work will be shared with the following relevant CIE Division 2 Technical Committees:

- TC 2-81 Update of CIE 065:1985 (Absolute radiometers)
- TC 2-90 LED Reference Spectrum for Photometer Calibration
- TC 2-96 Revision of ISO/CIE 19476:2014 Characterization of the Performance of Illuminance Meters and Luminance Meters

The project will contribute to preparing updates to and replacements for CIE 250:2022 Spectroradiometric measurement of optical radiation sources and CIE 127:2007 Measurement of LEDs.

Many European regulations and standards refer directly or indirectly to spectrally resolved absolute measurements of optical radiation power, including European standard EN 14255 "Measurement and assessment of personal exposures to incoherent optical radiation" parts 1 to 4, and EN/IEC 62471:2008/CIE S 009:2002 - Photobiological safety of lamps and lamp systems". These standards require spectrally resolved measurements in the spectral range of about 200 nm to 2000 nm, both in the laboratory and onsite. Improved calibration uncertainty of spectrally resolved equipment will be available as a result of this project.

#### *Longer-term economic, social and environmental impacts*

The European photonics industry has grown from €76 billion in 2015 to €103 billion in 2019 with a growth rate of 7 % per year and a share of 16 % of the global market. Besides components and materials, the major application segments are photonics for consumers' IT, medicine and biology, environment, lighting and energy, and Industry 4.0, indicating its direct involvement in the challenges for the future. The largest segment in Europe, photonics systems for industry – was worth €19.2 billion in 2019 and accounted for a 40 % share in the global markets. Europe is well positioned in the field of photonics systems for production, i.e. industrial laser systems, semiconductor manufacturing and machine vision.

Photonics21 is one of the European Technology Platforms supporting the EU-defined KETs and has more than 3000 members from the photonic industry, research institutes, academia, and public service. Established contacts between the project consortium and Photonics21 will simplify the transfer of knowledge about project outputs to this important technology platform. Successful project outputs will meet Photonics21 WG5 strategic roadmap targets for the period 2021 to 2027, where they request "maintenance-free, self-calibrating sensors, high quantum efficiency" as both a technology challenge and a research and innovation challenge for optimised value; a need outlined also by Quantum Flagship. The microelectronics-based primary standard detector is suitable for miniaturisation and may be integrated with the evolving technology of Photonic Integrated Circuits (PICs) in the future.

The incoming and outgoing radiation measurements are key components of Earth Energy Imbalance (EEI). Using PQEDs as a reference in lab-based calibrations or in future spaceflights could improve future EEI measurements.

Healthcare systems are under increasing pressure due to the aging population. High accuracy radiometric measurements are needed for faster and less invasive techniques in medical diagnostic instruments. This ensures cost efficiency, for example, by enabling analyses of blood samples at the point of care and the fast return of results while the patient is present.

## List of publications

-

This list is also available here: <https://www.euramet.org/repository/research-publications-repository-link/>

Project start date and duration:		01 June 2023, 36 months	
Coordinator: Jarle Gran, JV		Tel: +4764 848445; +4790 794154	E-mail: jag@justervesenet.no
Project website address: -			
Internal Beneficiaries: 1. JV, Norway 2. Aalto, Finland 3. CMI, Czechia 4. CSIC, Spain 5. INRIM, Italy 6. Metrosert, Estonia 7. PTB, Germany 8. RISE, Sweden 9. TUBITAK, Türkiye	External Beneficiaries: 10. IFE, Norway 11. SINTEF, Norway 12. USN, Norway	Unfunded Beneficiaries: 13. LMT, Germany	
Associated Partners: 14. SFI Davos, Switzerland			